NYC 9/11 Public Portal Document

AMA Analytical Services, Inc.

Fiber Count Sheet For Transmission Electron Microscopy

Date of Analysis: 9/24/01 | AMA Sample #: 0177773 | Filter Type/Size: MCE 25 mm PC ____ mm

Client: IT Corp. | Sample #: ET 0925-P3 | Filter Porosity: NP µm, Lot #: NP

Chain of Custody #: 87464 | Grid Acceptable (Y) N | Orientation of Letter F: ↖ ↗

Analyst: GECarey | Grid Storage Box #: 01387 | Working Mag: 15 K | L.O.D./Sensivity: ____ s/mm² 0. ____ s/cc

Signed: ____ | Row #: 10 Grid: A/B/C | Accel. Voltage: 100 kV | Volume of Air Sampled: 454 L

Method / type (circle): EPA Level II (AHERA) Other: ____ (Air) Water Wipe | Grid Openings to observe: 8

Grid Opening # Location	Structure #	Structure Type	Length & Width (In Microns)	SAED # Rows	Spectrum #	Elements Identified	Fiber Negative #	SAED Negative #	Magnification Or Camera Length
1/Y1	1	M	2.05 : 0.05	NP	—				K CM
2/K9	2	M	7.5 : 0.11	Pos	Ch, Cl				K CM
	3	M	0.5 : 0.07	Pos	Ch, Cl				K CM
3/J2	4	M	0.75 : 0.06	Pos	Ch, Cl				K CM
	5	M	1.05 : 0.15	Pos	Ch				K CM
4/N8	6	F	3.45 : 0.09	Pos	Ch				K CM
	7	M	1.45 : 0.05	Pos	Ch				K CM
5/E2 X	8	B	12.8 : 0.4	Pos	Ch				K CM
	9	M	1.75 : 0.17	Pos	Ch				K CM
	10	M	1.2 : 0.12	Pos	Ch				K CM
6/Q9 X	11	F	2.6 : 0.07	Pos	Ch				K CM
	12	F	1.0 : 0.07	Pos	Ch				K CM
	13	M	1.5 : 0.10	Pos	Ch				K CM
	14	F	2.3 : 0.07	Pos	Ch				K CM
	15	F	1.25 : 0.07	Pos	Ch				K CM

Legend: NSD = No Structures Detected F = Fiber B = Bundle C = Cluster M = Matrix UTO = Unable to Obtain ☒ = 2nd Grid

Total # of Grid Openings Observed 8 = 0.088 mm² # of Structures Counted: 19

Asbestos Type: Ch+Cl 15 s <5µm + 3 s ≥5µm = ____ s/mm² = ____ s/cc

Asbestos Type: ____ ____ s <5µm + ____ s ≥5µm = ____ s/mm² = ____ s/cc

Non-Asbestos Type: ____ 1 s <5µm + ____ s ≥5µm = ____ s/mm² = ____ s/cc

Total asbestos structures counted = 18 s = 205 s/mm² 0.1731 s/cc

SAED Calibration Date: 9/21/01 | Mean Grid Opening Area is 0.0 110 mm² | Notes: ~48≥50? particulate - occluded

EDXA Calibration Date: / / | Spectra Disk Identification #: ____ | Result may be biased - Note on Report

Mag. Calibration Date: 9/21/01 | EM Serial Number is: 156120-35 | 452-0 06.003189

CITY_21MC100_006788175

NYC-WTC_000429762