NYC 9/11 Public Portal Document

ATC ASSOCIATES, INC
104 E 25th Street, 10th Floor
New York, NY 10010
Tel. 212-353-8280
Fax 212-353-8306

REPORT DATE:	07-15-02
CLIENT:	Taylor Environmental Group, Inc.
PROJECT NAME:	SCA/Stuyvesant HS 345 Chambers St., 10th Floor
SAMPLED BY:	Client
SAMPLE DATE	07-14-02
DATE RECEIVED:	07-15-02
DATE ANALYZED:	07-15-02
SAMPLE MEDIA:	Air Samples
ANALYSIS REQUIRED:	Method NIOSH 7082

ANALYSIS RESULTS

SAMPLE ID	BATCH NO.	SAMPLE DESCRIPTION	LEAD CONCENTRATION ($\mu g/m^3$)	DETECTION LIMIT ($\mu g/m^3$)
LA-01	20367	Entrance of Room 1001	<4.2	4.2
LA-02	20367	Entrance of Room 1007	<4.2	4.2
LA-03	20367	Entrance of Room 1009	<1.8	1.8
LA-04	20367	Entrance of Room 1017	<1.8	1.8
LA-05	20367	Entrance of Room 1033	<1.9	1.9
LA-06	20367	Entrance of Room 1025	<3.8	3.8
LA-07	20367	Entrance of Room 1039	<3.6	3.6
LA-08	20367	Field Blank	<1.5 µg	1.5 µg
LA-09	20367	Field Blank	<1.5 µg	1.5 µg

NOTE 1:
THE REPORTING LIMIT (RL) IS 3.0 µg
NOTE 2:
At the present time there are no regulatory levels or requirements for indoor airborne lead in area samples. However, OSHA has established an action level of 30 $\mu g/m^3$ and a permissible exposure level (PEL) of 50 $\mu g/m^3$ for personal samples.
NOTE 3:
Blank units are in µg