NYC 9/11 Public Portal Document

FROM:
Theodorellys, Penny

TO:
Atteloney, Leonard

DATE:
08/23/2002 03:32:49 PM

SUBJECT:
TEM reporting results

---

p3 Asbestos Air Sample Collection and Analysis

The part in parentheses has been added. Is it okay?

Therefore, the laboratory shall report PCM, TEM (# structures <5 microns, # structures >5 microns, air concentration for structures <5 microns, air concentration for structures >5 microns, and structures per mm2) and PCMe values for each sample.

NYC-WTC_000433546