NYC 9/11 Public Portal Document

TRC
Customer-Focused Solutions

1430 Broadway, 10th Fl. New York, New York 10018
Phone: (212) 221-7822 - Fax: (212) 221-7840

Airborne Asbestos Fiber Analysis by Transmission Electron Microscopy

Client	Lower Manhattan Development Corporation	Job Number	46600-0000-00001
Address	One Liberty Plaza, 20th Floor	Date Received	2/5/2005
	New York, NY 10006	Date Analyzed	2/5/2005
Collected By	K.DaRocha	Job Location	130 Liberty Street
Collected	02/05/2005		
Rotometer	Standard		

Lab ID / Field ID	Sample Location	Start/ Stop	Flow	Volume (liters)	Asbestos Type(s)	Asbestos Structures ≥ .5mm <5mm	Asbestos Structures ≥ 5mm	Non-Asbestos	Analytical Sensitivity (AS/cc)	Asbestos Concentration (AS/cc)	Asbestos Concentration (AS/mm2)
030502938-0001	WASHINGTON STREET/ ALBANY STREET	00:03 08:01	4	1912	None Detected	0	0	0	0.0042	<0.0042	<21.00
T-020505-188											
030502938-0002	ALBANY STREET/ GREENWICH STREET	00:06 08:03	4	1908	None Detected	0	0	0	0.0042	<0.0042	<21.00
T-020505-189											
030502938-0003	GREENWICH STREET	00:10 08:06	4	1880.2	None Detected	0	0	0	0.0043	<0.0043	<21.00
T-020505-190											
030502938-0004	GREENWICH STREET/ LIBERTY STREET	00:12 08:09	4	1884.2	None Detected	0	0	0	0.0043	<0.0043	<21.00
T-020505-191											
030502938-0005	WASHINGTON STREET/ CEDAR STREET	00:17 08:13	4	1904	None Detected	0	0	0	0.0042	<0.0042	<21.00
T-020505-192											

Samples analyzed by EMSL Analytical, Inc.; NY ELAP ID# 10872 and 11506

CITY_21MC100_003228390

NYC-WTC_000442995

NYC 9/11 Public Portal Document

TRC
Customer-Focused Solutions

1430 Broadway, 10th Fl. New York, New York 10018
Phone: (212) 221-7822 - Fax: (212) 221-7840

Airborne Asbestos Fiber Analysis
by Transmission Electron Microscopy

Client	Lower Manhattan Development Corporation	Job Number	46600-0000-00001
Address	One Liberty Plaza, 20th Floor	Date Received	2/5/2005
	New York, NY 10006	Date Analyzed	2/5/2005
Collected By	K. DaRocha	Date Analyzed	130 Liberty Street
Collected	02/05/2005		
Rotometer	Standard		

Lab ID / Field ID	Sample Location	Start/ Stop	Flow	Volume (liters)	Asbestos Type(s)	Asbestos Structures ≥ .5mm <5mm	Asbestos Structures ≥ 5mm	Non-Asbestos	Analytical Sensitivity (AS/cc)	Asbestos Concentration (AS/cc)	Asbestos Concentration (AS/mm2)
030502939-0001 T-020505-194	WASHINGTON STREET/ ALBANY STREET	08:02 16:02	4	1920	None Detected	0	0	0	0.0042	<0.0042	<21.00
030502939-0002 T-020505-195	ALBANY STREET/ GREENWICH STREET	08:04 16:05	4	1924	None Detected	0	0	0	0.0042	<0.0042	<21.00
030502939-0003 T-020505-196	GREENWICH STREET	08:07 16:08	4	1876	None Detected	0	0	0	0.0043	<0.0043	<21.00
030502939-0004 T-020505-197	GREENWICH STREET/ LIBERTY STREET	08:11 16:12	4	1924	None Detected	0	0	0	0.0042	<0.0042	<21.00
030502939-0005 T-020505-198	WASHINGTON STREET/ CEDAR STREET	08:14 16:15	4	1900	None Detected	0	0	0	0.0042	<0.0042	<21.00

Samples analyzed by EMSL Analytical, Inc.; NY ELAP ID# 10872 and 11506

CITY_21MC100_003228391

NYC-WTC_000442996

NYC 9/11 Public Portal Document

TRC
Customer-Focused Solutions

1430 Broadway, 10th Fl. New York, New York 10018
Phone: (212) 221-7822 - Fax: (212) 221-7840

Airborne Asbestos Fiber Analysis
by Transmission Electron Microscopy

Client	Lower Manhattan Development Corporation	Job Number	46600-0000-00001
Address	One Liberty Plaza, 20th Floor	Date Received	02/06/2005
	New York, NY 10006	Date Analyzed	02/06/2005
Collected By	K.DaRocha	Date Analyzed	130 Liberty Street
Collected	02/05/2005		
Rotometer	Standard		

Lab ID / Field ID	Sample Location	Start/ Stop	Flow	Volume (liters)	Asbestos Type(s)	Asbestos Structures ≥ .5mm <5mm	Asbestos Structures ≥ 5mm	Non-Asbestos	Analytical Sensitivity (AS/cc)	Asbestos Concentration (AS/cc)	Asbestos Concentration (AS/mm2)
030502977-0001 T-020505-200	WASHINGTON STREET/ ALBANY STREET	16:02 00:02	4	1920.0	None Detected	0	0	0	0.0042	<0.0042	<21.00
030502977-0002 T-020505-201	ALBANY STREET/ GREENWICH STREET	16:05 00:03	4	1912.0	None Detected	0	0	0	0.0042	<0.0042	<21.00
030502977-0003 T-020505-202	GREENWICH STREET	16:08 00:07	4	1892.1	None Detected	0	0	0	0.0042	<0.0042	<21.00
030502977-0004 T-020505-203	GREENWICH STREET/ LIBERTY STREET	16:12 00:10	4	1912.0	None Detected	0	0	0	0.0042	<0.0042	<21.00
030502977-0005 T-020505-204	WASHINGTON STREET/ CEDAR STREET	16:15 00:14	4	1892.1	None Detected	0	0	0	0.0042	<0.0042	<21.00

Samples analyzed by EMSL Analytical, Inc.; NY ELAP ID# 10872 and 11506

CITY_21MC100_003228392

NYC-WTC_000442997