NYC-WTC_000445959

SUBJECT TO PROTECTIVE ORDER NO. 2, SEPT. 2008, CONFIDENTIAL, CITY_21MC100_00785006

NYC 9/11 Public Portal Document

AMBIENT GROUP, INC.

10 Morris Avenue, Glen Cove, NY 11542
Phone: (516) 671-2370 - Fax: (516) 609-8271
NY ELAP ID#10865

Airborne Asbestos Fiber Analysis by Transmission Electron Microscopy

Client: Lower Manhattan Development Corp.
Address: One Liberty Plaza, 20th Floor
New York, NY 10006

Collected By: Rotimi Adeola
Collected: 12/24/2004 6:06
Rotometer: Standard

Job Number: 2043
Date Received: 12/27/2004
Date Analyzed: 01/03/2005
Job Location: 130 Liberty Street

Lab ID / Field ID	Sample Location	Start/ Stop	Flow	Volume (liters)	Asbestos Type(s)	Asbestos Structures ≥ 5µm <5µm	Asbestos Structures ≥ 5µm	Non-Asbestos	Analytical Sensitivity (AS/cc)	Asbestos Concentration (AS/cc)	Asbestos Concentration (AS/mm2)
0013610-001 27678-7	Washington Street/ Albany Street	6:06 14:06	4	1920	None Detected	0	0	0	0.0008	<0.0008	<3.98
0013610-002 27678-8	Albany Street/ Greenwich Street	6:13 14:09	4.5	2142	None Detected	0	0	0	0.0008	<0.0008	<4.43
0013610-003 27678-9	Greenwich Street	6:21 14:11	4.25	1997.5	None Detected	0	0	0	0.0008	<0.0008	<3.98
0013610-004 27678-10	Liberty Street	6:30 14:15	4.75	2208.75	None Detected	0	0	0	0.0008	<0.0008	<4.69
0013610-005 27678-11	Washington Street/ Cedar Street	6:44 14:18	4.25	1929.5	None Detected	0	0	0	0.0008	<0.0008	<3.98

SIGNED FOR AMBIENT GROUP, INC. BY:

Disclaimers: The laboratory is not responsible for fibers counted in fibers/cc which is dependent on volume collected by non-laboratory personnel.
This report may not be duplicated in part without written permission by Ambient Group, Inc. This report must not be used to claim product endorsement by NVLAP or any agency of the U.S. Government.
This report relates only to the samples listed above.

NYC-WTC_000445960

NYC 9/11 Public Portal Document

SUBJECT TO PROTECTIVE ORDER NO. 2, SEPT. 2008 CONFIDENTIAL,

CITY_21MC100_007850065

AMBIENT GROUP, INC.

10 Morris Avenue, Glen Cove, NY 11542
Phone: (516) 671-2370 - Fax: (516) 609-8271
NY ELAP ID#10865

Airborne Asbestos Fiber Analysis by Transmission Electron Microscopy

Client: Lower Manhattan Development Corp.
Address: One Liberty Plaza, 20th Floor
New York, NY 10006

Collected By: Dennis Ikhide
Collected: 12/24/2004 14:06
Rotometer: Standard

Job Number: 2043
Date Received: 12/27/2004
Date Analyzed: 01/03/2005
Job Location: 130 Liberty Street

Lab ID / Field ID	Sample Location	Start/ Stop	Flow	Volume (liters)	Asbestos Type(s)	Asbestos Structures ≥.5µm <5µm	Asbestos Structures ≥ 5µm	Non-Asbestos	Analytical Sensitivity (AS/cc)	Asbestos Concentration (AS/cc)	Asbestos Concentration (AS/mm2)
0013611-001	Washington Street/ Albany Street	14:06 22:06	4	1920	None Detected	0	0	0	0.0008	<0.0008	<3.98
27681-7											
0013611-002	Albany Street/ Greenwich Street	14:09 22:10	4	1924	None Detected	0	0	0	0.0008	<0.0008	<3.98
27681-8											
0013611-003	Greenwich Street	14:11 22:13	4	1928	None Detected	0	0	0	0.0008	<0.0008	<3.98
27681-9											
0013611-004	Liberty Street	14:15 22:18	4	1932	None Detected	0	0	0	0.0008	<0.0008	<3.98
27681-10											
0013611-005	Washington Street/ Cedar Street	14:18 22:22	4	1936	None Detected	0	0	0	0.0008	<0.0008	<3.98
27681-11											

SIGNED FOR AMBIENT GROUP, INC. BY:

Disclaimers: The laboratory is not responsible for fibers counted in fibers/cc which is dependent on volume collected by non-laboratory personnel.
This report may not be duplicated in part without written permission by Ambient Group, Inc. This report must not be used to claim product endorsement by NVLAP or any agency of the U.S. Government.
This report relates only to the samples listed above.

NYC-WTC_000445961

NYC 9/11 Public Portal Document

SUBJECT TO PROTECTIVE ORDER NO. 2, SEPT. 2008 CONFIDENTIAL,

CITY_21MC100_007850066

AMBIENT GROUP, INC.

10 Morris Avenue, Glen Cove, NY 11542
Phone: (516) 671-2370 - Fax: (516) 609-8271
NY ELAP ID#10865

Airborne Asbestos Fiber Analysis by Transmission Electron Microscopy

Client: Lower Manhattan Development Corp.
Address: One Liberty Plaza, 20th Floor
New York, NY 10006

Collected By: Ike Onyenyili
Collected: 12/24/2004 22:06
Rotometer: Standard

Job Number: 2043
Date Received: 12/27/2004
Date Analyzed: 01/03/2005
Job Location: 130 Liberty Street

Lab ID / Field ID	Sample Location	Start/ Stop	Flow	Volume (liters)	Asbestos Type(s)	Asbestos Structures ≥ 5µm <5µm	Asbestos Structures ≥ 5µm	Non-Asbestos	Analytical Sensitivity (AS/cc)	Asbestos Concentration (AS/cc)	Asbestos Concentration (AS/mm2)
0013612-001 27683-7	Washington Street/ Albany Street	22:06 6:05	4	1916	None Detected	0	0	0	0.0008	<0.0008	<3.98
0013612-002 27683-8	Albany Street/ Greenwich Street	22:10 6:11	4	1924	None Detected	0	0	0	0.0008	<0.0008	<3.98
0013612-003 27683-9	Greenwich Street	22:13 6:19	4	1944	None Detected	0	0	0	0.0008	<0.0008	<3.98
0013612-004 27683-10	Liberty Street	22:18 6:27	4	1956	None Detected	0	0	0	0.0008	<0.0008	<3.98
0013612-005 27683-11	Washington Street/ Cedar Street	22:22 6:45	4	2012	None Detected	0	0	0	0.0008	<0.0008	<4.19

Average Filter Loading: <3.98

SIGNED FOR AMBIENT GROUP, INC. BY:

Disclaimers: The laboratory is not responsible for fibers counted in fibers/cc which is dependent on volume collected by non-laboratory personnel.
This report may not be duplicated in part without written permission by Ambient Group, Inc. This report must not be used to claim product endorsement by NVLAP or any agency of the U.S. Government.
This report relates only to the samples listed above.